

# Clinical Laboratory of San Bernardino, Inc.



**Executive Environmental**  
310 E. Foothill Blvd., Suite 200  
Arcadia CA, 91006

Project: Water Analysis  
Sub Project: 22-Z0043-0082.6  
Project Manager: Daniel H. Ginsborg

Work Order: 22L1739  
Received: 12/14/22 17:06  
Reported: 01/04/23

**304370407 A**

**22L1739-01 (Water)**

**Sample Date:** 12/08/22 6:19

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**

EPA 200.8

1.8

ug/L

1.0

12/29/22

12/29/22

2253089

**304370407 B**

**22L1739-02 (Water)**

**Sample Date:** 12/08/22 6:20

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**

EPA 200.8

1.2

ug/L

1.0

12/29/22

12/29/22

2253089

**304370407 C**

**22L1739-03 (Water)**

**Sample Date:** 12/08/22 6:21

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**

EPA 200.8

0.47

ug/L

1.0

12/29/22

12/29/22

2253089

J

**304370407 D**

**22L1739-04 (Water)**

**Sample Date:** 12/08/22 6:22

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**

EPA 200.8

0.85

ug/L

1.0

12/29/22

12/29/22

2253089

J

**304370407 E**

**22L1739-05 (Water)**

**Sample Date:** 12/08/22 6:24

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**

EPA 200.8

0.37

ug/L

1.0

12/29/22

12/29/22

2253089

J

**304370407 F**

**22L1739-06 (Water)**

**Sample Date:** 12/08/22 6:25

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**

EPA 200.8

1.1

ug/L

1.0

12/29/22

12/29/22

2253089

**304370407 G**

**22L1739-07 (Water)**

**Sample Date:** 12/08/22 6:35

**Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

# Clinical Laboratory of San Bernardino, Inc.



<b>Executive Environmental</b> 310 E. Foothill Blvd., Suite 200 Arcadia CA, 91006	Project: Water Analysis Sub Project: 22-Z0043-0082.6 Project Manager: Daniel H. Ginsborg	Work Order: 22L1739 Received: 12/14/22 17:06 Reported: 01/04/23
---	--	---

**304370407 G** **22L1739-07 (Water)** **Sample Date:** 12/08/22 6:35 **Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**      EPA 200.8      **0.94**      ug/L      1.0           12/29/22      12/29/22      2253089      J

**304370407 H** **22L1739-08 (Water)** **Sample Date:** 12/08/22 6:36 **Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**      EPA 200.8      **0.70**      ug/L      1.0           12/29/22      12/29/22      2253089      J

**304370407 A30** **22L1739-09 (Water)** **Sample Date:** 12/08/22 6:39 **Sampler:** Matt Barna

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

**Metals**

**Lead (Low Level)**      EPA 200.8      **0.91**      ug/L      1.0           12/29/22      12/29/22      2253089      J

J      Detected below the Reporting Limit; reported concentration is estimated; (J-Flag)

ND      Analyte NOT DETECTED at or above the reporting limit

**Jeanette Hernandez**  
Project Manager

# Clinical Laboratory of San Bernardino, Inc.



**Executive Environmental**  
310 E. Foothill Blvd., Suite 200  
Arcadia CA, 91006

Project: Water Analysis  
Sub Project: 22-Z0043-0082.6  
Project Manager: Daniel H. Ginsborg

Work Order: 22L1739  
Received: 12/14/22 17:06  
Reported: 01/04/23

## Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
<b>Batch 2253089      Analyst: AP1</b>										
<b>Blank (2253089-BLK1)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK2)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK3)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK4)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	0.14	1.0	ug/L							J
<b>Blank (2253089-BLK5)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK6)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK7)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK8)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>Blank (2253089-BLK9)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	ND	1.0	ug/L							
<b>LCS (2253089-BS1)</b> Prepared & Analyzed: 12/29/22										
Lead (Low Level)	49.4	1.0	ug/L	50		99	85-115			

# Clinical Laboratory of San Bernardino, Inc.



<b>Executive Environmental</b> 310 E. Foothill Blvd., Suite 200 Arcadia CA, 91006	Project: Water Analysis Sub Project: 22-Z0043-0082.6 Project Manager: Daniel H. Ginsborg	Work Order: 22L1739 Received: 12/14/22 17:06 Reported: 01/04/23
---	--	---

## Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
<b>Batch 2253089</b>		<b>Analyst: AP1</b>								
<b>LCS (2253089-BS2)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	50.2	1.0	ug/L	50		100	85-115			
<b>LCS (2253089-BS3)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	52.4	1.0	ug/L	50		105	85-115			
<b>LCS (2253089-BS4)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	53.5	1.0	ug/L	50		107	85-115			
<b>LCS (2253089-BS5)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	53.8	1.0	ug/L	50		108	85-115			
<b>LCS (2253089-BS6)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	53.8	1.0	ug/L	50		108	85-115			
<b>LCS (2253089-BS7)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	54.0	1.0	ug/L	50		108	85-115			
<b>LCS (2253089-BS8)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	54.3	1.0	ug/L	50		109	85-115			
<b>LCS (2253089-BS9)</b>				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	54.6	1.0	ug/L	50		109	85-115			
<b>Matrix Spike (2253089-MS1)</b>				Source: 22L1440-01		Prepared & Analyzed: 12/29/22				
Lead (Low Level)	51.4	1.0	ug/L	50	0.825	101	70-130			
<b>Matrix Spike (2253089-MS2)</b>				Source: 22L1440-11		Prepared & Analyzed: 12/29/22				
Lead (Low Level)	50.4	1.0	ug/L	50	0.266	100	70-130			

# Clinical Laboratory of San Bernardino, Inc.



<b>Executive Environmental</b> 310 E. Foothill Blvd., Suite 200 Arcadia CA, 91006	Project: Water Analysis Sub Project: 22-Z0043-0082.6 Project Manager: Daniel H. Ginsborg	Work Order: 22L1739 Received: 12/14/22 17:06 Reported: 01/04/23
---	--	---

## Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-----------

**Batch 2253089 Analyst: AP1**

<b>Matrix Spike (2253089-MS3)</b>	<b>Source: 22L1734-03</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	52.3	1.0 ug/L		50	0.119	104	70-130			
<b>Matrix Spike (2253089-MS4)</b>	<b>Source: 22L1737-01</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	58.1	1.0 ug/L		50	4.47	107	70-130			
<b>Matrix Spike (2253089-MS5)</b>	<b>Source: 22L1738-06</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	59.3	1.0 ug/L		50	5.19	108	70-130			
<b>Matrix Spike (2253089-MS6)</b>	<b>Source: 22L1739-09</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	55.0	1.0 ug/L		50	0.913	108	70-130			
<b>Matrix Spike (2253089-MS7)</b>	<b>Source: 22L1741-01</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	54.9	1.0 ug/L		50	0.559	109	70-130			
<b>Matrix Spike Dup (2253089-MSD1)</b>	<b>Source: 22L1440-01</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	50.4	1.0 ug/L		50	0.825	99	70-130	2	30	
<b>Matrix Spike Dup (2253089-MSD2)</b>	<b>Source: 22L1440-11</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	50.4	1.0 ug/L		50	0.266	100	70-130	0.1	30	
<b>Matrix Spike Dup (2253089-MSD3)</b>	<b>Source: 22L1734-03</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	52.5	1.0 ug/L		50	0.119	105	70-130	0.4	30	
<b>Matrix Spike Dup (2253089-MSD4)</b>	<b>Source: 22L1737-01</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	57.4	1.0 ug/L		50	4.47	106	70-130	1	30	
<b>Matrix Spike Dup (2253089-MSD5)</b>	<b>Source: 22L1738-06</b>		<b>Prepared &amp; Analyzed: 12/29/22</b>							
Lead (Low Level)	58.6	1.0 ug/L		50	5.19	107	70-130	1	30	

# Clinical Laboratory of San Bernardino, Inc.



**Executive Environmental**  
310 E. Foothill Blvd., Suite 200  
Arcadia CA, 91006

Project: Water Analysis  
Sub Project: 22-Z0043-0082.6  
Project Manager: Daniel H. Ginsborg

Work Order: 22L1739  
Received: 12/14/22 17:06  
Reported: 01/04/23

## Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-----------

**Batch 2253089**

**Analyst: AP1**

**Matrix Spike Dup (2253089-MSD6)**

Source: 22L1739-09

Prepared & Analyzed: 12/29/22

Lead (Low Level)	56.4	1.0 ug/L		50	0.913	111	70-130	2	30	
------------------	------	----------	--	----	-------	-----	--------	---	----	--

**Matrix Spike Dup (2253089-MSD7)**

Source: 22L1741-01

Prepared & Analyzed: 12/29/22

Lead (Low Level)	55.2	1.0 ug/L		50	0.559	109	70-130	0.5	30	
------------------	------	----------	--	----	-------	-----	--------	-----	----	--

J Detected below the Reporting Limit; reported concentration is estimated; (J-Flag)

ND Analyte NOT DETECTED at or above the reporting limit

**Jeanette Hernandez**

**Project Manager**