

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental
310 E. Foothill Blvd., Suite 200
Arcadia CA, 91006

Project: Water Analysis
Sub Project: 22-Z0043-0082.4
Project Manager: Daniel H. Ginsborg

Work Order: 22L1733
Received: 12/14/22 17:06
Reported: 01/04/23

304371126 A

22L1733-01 (Water)

Sample Date: 12/09/22 6:09

Sampler: Rhys Kuzmic

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

Metals

Lead (Low Level)

EPA 200.8

0.42

ug/L

1.0

12/29/22

12/29/22

2253089

J

304371126 B

22L1733-02 (Water)

Sample Date: 12/09/22 6:10

Sampler: Rhys Kuzmic

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

Metals

Lead (Low Level)

EPA 200.8

0.38

ug/L

1.0

12/29/22

12/29/22

2253089

J

304371126 A30

22L1733-03 (Water)

Sample Date: 12/09/22 6:12

Sampler: Rhys Kuzmic

Analyte	Method	Result	Units	Rep. Limit	MCL	Prepared	Analyzed	Batch	Qualifier
---------	--------	--------	-------	------------	-----	----------	----------	-------	-----------

Metals

Lead (Low Level)

EPA 200.8

0.44

ug/L

1.0

12/29/22

12/29/22

2253089

J

J Detected below the Reporting Limit; reported concentration is estimated; (J-Flag)

ND Analyte NOT DETECTED at or above the reporting limit

Jeanette Hernandez
Project Manager



Executive Environmental 310 E. Foothill Blvd., Suite 200 Arcadia CA, 91006	Project: Water Analysis Sub Project: 22-Z0043-0082.4 Project Manager: Daniel H. Ginsborg	Work Order: 22L1733 Received: 12/14/22 17:06 Reported: 01/04/23
---	--	---

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
Batch 2253089		Analyst: AP1								
Blank (2253089-BLK1)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK2)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK3)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK4)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	0.14	1.0	ug/L							J
Blank (2253089-BLK5)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK6)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK7)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK8)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
Blank (2253089-BLK9)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	ND	1.0	ug/L							
LCS (2253089-BS1)	Prepared & Analyzed: 12/29/22									
Lead (Low Level)	49.4	1.0	ug/L	50		99	85-115			

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental 310 E. Foothill Blvd., Suite 200 Arcadia CA, 91006	Project: Water Analysis Sub Project: 22-Z0043-0082.4 Project Manager: Daniel H. Ginsborg	Work Order: 22L1733 Received: 12/14/22 17:06 Reported: 01/04/23
---	--	---

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
Batch 2253089 Analyst: API										
LCS (2253089-BS2)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	50.2	1.0	ug/L	50		100	85-115			
LCS (2253089-BS3)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	52.4	1.0	ug/L	50		105	85-115			
LCS (2253089-BS4)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	53.5	1.0	ug/L	50		107	85-115			
LCS (2253089-BS5)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	53.8	1.0	ug/L	50		108	85-115			
LCS (2253089-BS6)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	53.8	1.0	ug/L	50		108	85-115			
LCS (2253089-BS7)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	54.0	1.0	ug/L	50		108	85-115			
LCS (2253089-BS8)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	54.3	1.0	ug/L	50		109	85-115			
LCS (2253089-BS9)				Prepared & Analyzed: 12/29/22						
Lead (Low Level)	54.6	1.0	ug/L	50		109	85-115			
Matrix Spike (2253089-MS1)				Source: 22L1440-01		Prepared & Analyzed: 12/29/22				
Lead (Low Level)	51.4	1.0	ug/L	50	0.825	101	70-130			
Matrix Spike (2253089-MS2)				Source: 22L1440-11		Prepared & Analyzed: 12/29/22				
Lead (Low Level)	50.4	1.0	ug/L	50	0.266	100	70-130			

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental
310 E. Foothill Blvd., Suite 200
Arcadia CA, 91006

Project: Water Analysis
Sub Project: 22-Z0043-0082.4
Project Manager: Daniel H. Ginsborg

Work Order: 22L1733
Received: 12/14/22 17:06
Reported: 01/04/23

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
Batch 2253089 Analyst: AP1										
Matrix Spike (2253089-MS3) Source: 22L1734-03 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	52.3	1.0 ug/L		50	0.119	104	70-130			
Matrix Spike (2253089-MS4) Source: 22L1737-01 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	58.1	1.0 ug/L		50	4.47	107	70-130			
Matrix Spike (2253089-MS5) Source: 22L1738-06 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	59.3	1.0 ug/L		50	5.19	108	70-130			
Matrix Spike (2253089-MS6) Source: 22L1739-09 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	55.0	1.0 ug/L		50	0.913	108	70-130			
Matrix Spike (2253089-MS7) Source: 22L1741-01 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	54.9	1.0 ug/L		50	0.559	109	70-130			
Matrix Spike Dup (2253089-MSD1) Source: 22L1440-01 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	50.4	1.0 ug/L		50	0.825	99	70-130	2	30	
Matrix Spike Dup (2253089-MSD2) Source: 22L1440-11 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	50.4	1.0 ug/L		50	0.266	100	70-130	0.1	30	
Matrix Spike Dup (2253089-MSD3) Source: 22L1734-03 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	52.5	1.0 ug/L		50	0.119	105	70-130	0.4	30	
Matrix Spike Dup (2253089-MSD4) Source: 22L1737-01 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	57.4	1.0 ug/L		50	4.47	106	70-130	1	30	
Matrix Spike Dup (2253089-MSD5) Source: 22L1738-06 Prepared & Analyzed: 12/29/22										
Lead (Low Level)	58.6	1.0 ug/L		50	5.19	107	70-130	1	30	

Clinical Laboratory of San Bernardino, Inc.



Executive Environmental 310 E. Foothill Blvd., Suite 200 Arcadia CA, 91006	Project: Water Analysis	Work Order: 22L1733
	Sub Project: 22-Z0043-0082.4	Received: 12/14/22 17:06
	Project Manager: Daniel H. Ginsborg	Reported: 01/04/23

Metals - Quality Control

Analyte	Result	Reporting Limit	Units	Spike Level	Source Result	%Rec	%Rec Limits	RPD	RPD Limit	Qualifier
---------	--------	-----------------	-------	-------------	---------------	------	-------------	-----	-----------	-----------

Batch 2253089

Analyst: AP1

Matrix Spike Dup (2253089-MSD6)	Source: 22L1739-09	Prepared & Analyzed: 12/29/22							
Lead (Low Level)	56.4	1.0 ug/L	50	0.913	111	70-130	2	30	
Matrix Spike Dup (2253089-MSD7)	Source: 22L1741-01	Prepared & Analyzed: 12/29/22							
Lead (Low Level)	55.2	1.0 ug/L	50	0.559	109	70-130	0.5	30	

J Detected below the Reporting Limit; reported concentration is estimated; (J-Flag)

ND Analyte NOT DETECTED at or above the reporting limit

Jeanette Hernandez

Project Manager